

Overview

Useful For

Investigating an undetectable total complement

Confirming hereditary angioedema (with low C1 inhibitor)

Assessing disease activity in systemic lupus erythematosus, proliferative glomerulonephritis, rheumatoid arthritis, and autoimmune hemolytic anemia

Method Name

Nephelometry

NY State Available

Yes

Specimen

Specimen Type

Serum

Specimen Required

Supplies: Sarstedt Aliquot Tube 5 mL (T914)

Collection Container/Tube:

Preferred: Serum gel

Acceptable: Red top

Submission Container/Tube: Plastic vial

Specimen Volume: 1 mL Serum

Collection Instructions: Centrifuge and aliquot serum into a plastic vial.

Forms

If not ordering electronically, complete, print, and send 1 of the following forms with the specimen:

[-Kidney Transplant Test Request](#)

[-Renal Diagnostics Test Request](#) (T830)

Specimen Minimum Volume

Serum: 0.5 mL

Reject Due To

Gross hemolysis	OK
-----------------	----

Gross lipemia	Reject
Gross icterus	OK

Specimen Stability Information

Specimen Type	Temperature	Time	Special Container
Serum	Refrigerated (preferred)	7 days	
	Ambient	72 hours	
	Frozen	28 days	

Clinical & Interpretive**Clinical Information**

The complement system is an integral part of the immune defenses. It can be activated via immune complexes (classic pathway) or by bacterial polysaccharides (alternative pathway). The classic complement pathway consists of recognition (C1q, C1r, C1s), activation (C2, C3, C4), and attack (C5, C6, C7, C8, C9) mechanisms with respect to their role in antibody-mediated cytolysis. C4 is one of the activation proteins of the classic pathway.

In the absence of C4, immune complexes will not be cleared by C3 activation peptides, but bacterial infections can still be defended via the alternative pathway.

Complement component 4 may be decreased in systemic lupus erythematosus, early glomerulonephritis, immune complex disease, cryoglobulinemia, hereditary angioedema, and congenital C4 deficiency.

Reference Values

14-40 mg/dL

Interpretation

Complement component 4 levels will be decreased in acquired autoimmune disorders, in the active phase of lupus erythematosus, and in rheumatoid arthritis.

An undetectable C4 level (with normal C3) suggests a congenital C4 deficiency.

Levels will be increased in patients with autoimmune hemolytic anemia.

Cautions

The results are dependent on appropriate specimen transport.

Quantitation of specific proteins by nephelometric means may not be possible in lipemic sera due to the extreme light scattering properties of the specimen. Turbidity and particles in the specimen may result in extraneous light scattering signals, resulting in variable specimen analysis.

Clinical Reference

1. Willrich MAV, Braun KMP, Moyer AM, Jeffrey DH, Frazer-Abel A. Complement testing in the clinical laboratory. Crit

Rev Clin Lab Sci. 2021;58(7):447-478. doi:10.1080/10408363.2021.19072972

2. Wong EKS, Kavanagh D. Diseases of complement dysregulation-an overview. Semin Immunopathol. 2018;40(1):49-64. doi:10.1007/s00281-017-0663-8

3. Prohaszka Z, Kirschfink M, Frazer-Abel A. Complement analysis in the era of targeted therapeutics. Mol Immunol. 2018;102:84-88. doi:10.1016/j.molimm.2018.06.001

4. Brodzski N, Frazer-Abel A, Grumach AS, et al. European Society for Immunodeficiencies (ESID) and European Reference Network on Rare Primary Immunodeficiency, Autoinflammatory and Autoimmune Diseases (ERN RITA) Complement Guideline: Deficiencies, Diagnosis, and Management. J Clin Immunol. 2020;40(4):576-591. doi:10.1007/s10875-020-00754-1

Performance

Method Description

In this Siemens Nephelometer II method, the light scattered onto the antigen-antibody complexes is measured. The intensity of the measured scattered light is proportional to the amount of antigen-antibody complexes in the sample under certain conditions. If the antibody volume is kept constant, the signal behaves proportionally to the antigen volume.

A reference curve is generated by a standard with a known antigen content on which the scattered light signals of the samples can be evaluated and calculated as an antigen concentration. Antigen-antibody complexes are formed when a sample containing antigen and the corresponding antiserum are put into a cuvette. A light beam is generated with a light-emitting diode, which is transmitted through the cuvette. The light is scattered onto the immuno-complexes that are present. Antigen and antibody are mixed in the initial measurement, but no complex is formed yet. An antigen-antibody complex is formed in the final measurement.

The result is calculated by subtracting value of the final measurement from the initial measurement. The distribution of intensity of the scattered light depends on the ratio of the particle size of the antigen-antibody complexes to the radiated wavelength. (Instruction manual: Siemens Nephelometer II Operations. Siemens, Inc; Version 2.4, 07/2019)

PDF Report

No

Day(s) Performed

Monday through Friday

Report Available

1 to 3 days

Specimen Retention Time

14 days

Performing Laboratory Location

Mayo Clinic Laboratories - Rochester Superior Drive

Fees & Codes**Fees**

- Authorized users can sign in to [Test Prices](#) for detailed fee information.
- Clients without access to Test Prices can contact [Customer Service](#) 24 hours a day, seven days a week.
- Prospective clients should contact their account representative. For assistance, contact [Customer Service](#).

Test Classification

This test has been cleared, approved, or is exempt by the US Food and Drug Administration and is used per manufacturer's instructions. Performance characteristics were verified by Mayo Clinic in a manner consistent with CLIA requirements.

CPT Code Information

86160

LOINC® Information

Test ID	Test Order Name	Order LOINC® Value
C4	Complement C4, S	4498-2

Result ID	Test Result Name	Result LOINC® Value
C4	Complement C4, S	4498-2